

SOLUTION CHANGE

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----


Date: 12/01/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:48
INTERNAL STANDARD	VERIFIED	17:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	INTERFERENCE	17:49

Interference Detected



SOLUTION CHANGE

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 12/01/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:52
INTERNAL STANDARD	VERIFIED	17:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:53
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:54
BLANK TEST	0.00	17:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:55
BLANK TEST	0.00	17:56

