

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 12/05/2021

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 15:28 |
| INTERNAL STANDARD       | VERIFIED | 15:28 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 15:29 |
| BLANK TEST              | 0.00     | 15:30 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 15:30 |
| BLANK TEST              | 0.00     | 15:31 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 15:31 |
| BLANK TEST              | 0.00     | 15:32 |

