

SUPERVISOR TEST

 CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 12/28/2021

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:12
INTERNAL STANDARD	VERIFIED	08:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:12
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15

