

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 02/08/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:24
INTERNAL STANDARD	VERIFIED	15:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:26
BLANK TEST	0.00	15:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:27
BLANK TEST	0.00	15:28

