

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 02/14/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 15:53 |
| INTERNAL STANDARD | VERIFIED | 15:53 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 15:54 |
| BLANK TEST | 0.00 | 15:54 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 15:55 |
| BLANK TEST | 0.00 | 15:56 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.10 | 15:56 |
| BLANK TEST | 0.00 | 15:57 |

