

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 02/21/2022

OPERATOR NAME:
JOHN R DEARMAN
Permit #: 36625
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:03
INTERNAL STANDARD	VERIFIED	13:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:04
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:05
BLANK TEST	0.00	13:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:06
BLANK TEST	0.00	13:07

