

SOLUTION CHANGE

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 03/01/2022

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:38
INTERNAL STANDARD	VERIFIED	00:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:40
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:41
BLANK TEST	0.00	00:42

