

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 04/03/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:41
INTERNAL STANDARD	VERIFIED	00:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:43
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:44
BLANK TEST	0.00	00:45

