

SUPERVISOR TEST

CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 04/24/2022

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:27
INTERNAL STANDARD	VERIFIED	15:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:28
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:29
BLANK TEST	0.00	15:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:30
BLANK TEST	0.00	15:31

