

SOLUTION CHANGE

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 05/03/2022

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:15
INTERNAL STANDARD	VERIFIED	00:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:18
BLANK TEST	0.00	00:19

