

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 05/27/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:46
INTERNAL STANDARD	VERIFIED	15:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:47
BLANK TEST	0.00	15:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:49
BLANK TEST	0.00	15:49

