

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 06/02/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:03
INTERNAL STANDARD	VERIFIED	17:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:06
BLANK TEST	0.00	17:06

