

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

ate: 06/07/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:56
INTERNAL STANDARD	VERIFIED	23:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:56
BLANK TEST	0.00	23:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:58
BLANK TEST	0.00	23:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:59
BLANK TEST	0.00	00:00

