

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 06/27/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:37
INTERNAL STANDARD	VERIFIED	09:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:38
BLANK TEST	0.00	09:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:39
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:40
BLANK TEST	0.00	09:41

