

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 07/03/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:55
INTERNAL STANDARD	VERIFIED	15:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:55
BLANK TEST	0.00	15:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:57
BLANK TEST	0.00	15:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:59

