

SUPERVISOR TEST

 CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 07/19/2022

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:28
INTERNAL STANDARD	VERIFIED	23:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:28
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:30
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:31
BLANK TEST	0.00	23:32

