

SOLUTION CHANGE

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----

Date: 08/01/2022

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:51
INTERNAL STANDARD	VERIFIED	07:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:51
BLANK TEST	0.00	07:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:53
BLANK TEST	0.00	07:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:54
BLANK TEST	0.00	07:55

