

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 08/07/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:41
INTERNAL STANDARD	VERIFIED	20:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:41
BLANK TEST	0.00	20:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:43
BLANK TEST	0.00	20:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:44
BLANK TEST	0.00	20:45

