

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 08/24/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:26
INTERNAL STANDARD	VERIFIED	23:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:27
BLANK TEST	0.00	23:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:28
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:29
BLANK TEST	0.00	23:30

