

SUPERVISOR TEST

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----

Date: 08/15/2022

OPERATOR NAME:  
JOHN R DEARMAN  
Permit #: 36625  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 09:45 |
| INTERNAL STANDARD       | VERIFIED | 09:45 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.10     | 09:46 |
| BLANK TEST              | 0.00     | 09:47 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 09:47 |
| BLANK TEST              | 0.00     | 09:48 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 09:48 |
| BLANK TEST              | 0.00     | 09:49 |

