

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 09/12/2022

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:26
INTERNAL STANDARD	VERIFIED	15:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:27
BLANK TEST	0.00	15:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:28
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:29
BLANK TEST	0.00	15:30

