

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 09/18/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:42
INTERNAL STANDARD	VERIFIED	15:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:45
BLANK TEST	0.00	15:46

