

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 10/18/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:23
INTERNAL STANDARD	VERIFIED	21:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:23
BLANK TEST	0.00	21:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:24
BLANK TEST	0.00	21:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:26
BLANK TEST	0.00	21:27

