

SUPERVISOR TEST

CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 10/24/2022

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:42
INTERNAL STANDARD	VERIFIED	15:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:42
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:45

