

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 11/23/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:28
INTERNAL STANDARD	VERIFIED	15:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:29
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:30
BLANK TEST	0.00	15:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:31
BLANK TEST	0.00	15:32

