

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 12/01/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:39
INTERNAL STANDARD	VERIFIED	18:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:39
BLANK TEST	0.00	18:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:40
BLANK TEST	0.00	18:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	18:41
BLANK TEST	0.00	18:42

