

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 12/19/2022

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:34
INTERNAL STANDARD	VERIFIED	12:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:34
BLANK TEST	0.00	12:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:35
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:36
BLANK TEST	0.00	12:37

