

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 12/29/2022

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:35
INTERNAL STANDARD	VERIFIED	16:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:36
BLANK TEST	0.00	16:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:37
BLANK TEST	0.00	16:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:38
BLANK TEST	0.00	16:39

