

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 01/01/2023

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:11
INTERNAL STANDARD	VERIFIED	16:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:11
BLANK TEST	0.00	16:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	16:12
BLANK TEST	0.00	16:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:13
BLANK TEST	0.00	16:14

