

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 02/01/2023

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:00
INTERNAL STANDARD	VERIFIED	00:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:00
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:01
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:03
BLANK TEST	0.00	00:04

