

SUPERVISOR TEST

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----

Date: 02/13/2023

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:41
INTERNAL STANDARD	VERIFIED	23:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:41
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:43
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45

