

SUPERVISOR TEST

CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 04/11/2023

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:23
INTERNAL STANDARD	VERIFIED	23:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:24
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:25
BLANK TEST	0.00	23:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:26
BLANK TEST	0.00	23:27

