

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 05/14/2023

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:26
INTERNAL STANDARD	VERIFIED	23:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:27
BLANK TEST	0.00	23:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:28
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:29
BLANK TEST	0.00	23:30

