

SOLUTION CHANGE

-----  
CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196  
-----

Date: 06/12/2023

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:10
INTERNAL STANDARD	VERIFIED	17:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:12
BLANK TEST	0.00	17:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	17:13
BLANK TEST	0.00	17:14

