

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 06/27/2023

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:20
INTERNAL STANDARD	VERIFIED	01:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:22
BLANK TEST	0.00	01:23

