

SUPERVISOR TEST

CARMEL TOWN
 POLICE DEPT
 DATAMASTER dmt: 100196

Date: 07/29/2023

OPERATOR NAME:
 STEPHEN M MCCABE
 Permit #: 41419
 Operator Agency/Dept: CARMEL TOWN PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:21
INTERNAL STANDARD	VERIFIED	00:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:21
BLANK TEST	0.00	00:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:23
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:24
BLANK TEST	0.00	00:25

