

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 08/03/2023

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:49
INTERNAL STANDARD	VERIFIED	23:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:50
BLANK TEST	0.00	23:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:51
BLANK TEST	0.00	23:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:52
BLANK TEST	0.00	23:53

