

SUPERVISOR TEST

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 09/24/2023

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:44
INTERNAL STANDARD	VERIFIED	23:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:45
BLANK TEST	0.00	23:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:46
BLANK TEST	0.00	23:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:47
BLANK TEST	0.00	23:48

