

SOLUTION CHANGE

CARMEL TOWN  
POLICE DEPT  
DATAMASTER dmt: 100196

Date: 10/02/2023

OPERATOR NAME:  
STEPHEN M MCCABE  
Permit #: 41419  
Operator Agency/Dept: CARMEL TOWN PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:01
INTERNAL STANDARD	VERIFIED	00:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:04

