

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 10/13/2023

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:39
INTERNAL STANDARD	VERIFIED	00:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43

