

SOLUTION CHANGE

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 11/05/2023

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:58
INTERNAL STANDARD	VERIFIED	23:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:58
BLANK TEST	0.00	23:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:00
BLANK TEST	0.00	00:01

