

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 01/12/2024

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:54
INTERNAL STANDARD	VERIFIED	00:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:58

