

SUPERVISOR TEST

CARMEL TOWN
POLICE DEPT
DATAMASTER dmt: 100196

Date: 01/19/2024

OPERATOR NAME:
STEPHEN M MCCABE
Permit #: 41419
Operator Agency/Dept: CARMEL TOWN PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:00
INTERNAL STANDARD	VERIFIED	14:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:00
BLANK TEST	0.00	14:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:02
BLANK TEST	0.00	14:03

